

DROP TEST CHLORIDE (1 drop = 2 ppm)

Instr. #5025

COMPONENTS:

1 x 5025	Instruction
1 x 91980	Sample Tube, Graduated (25 mL) w/ cap & orange dot, plastic
1 x R-0630-C	Chromate Indicator, 2 oz, DB
1 x R-06380-C	Phenolphthalein Indicator, 2 oz w/ orange cap, DB
1 x R-06860-C	Sulfuric Acid N, 2 oz w/ orange cap, DB
1 x R-0807-C	Silver Nitrate Reagent, 2 oz, DB

TO ORDER REPLACEMENT PARTS AND REAGENTS CALL TOLL-FREE
800-TEST KIT (800-837-8548).

PROCEDURE:

**CAREFULLY READ AND FOLLOW PRECAUTIONS ON REAGENT LABELS.
KEEP REAGENTS AWAY FROM CHILDREN.**

NOTE: When dispensing reagents from dropper bottles, **always** hold bottle in a vertical position.

Chloride Test

NOTE: When sulfite content of sample water to be tested exceeds 10 ppm, the sulfite should be oxidized to prevent interference in test. A 25 mL water sample is first adjusted to the appropriate pH, then 1 mL (or 25 drops) of R-0649 Hydrogen Peroxide Solution (sold separately) is added and thoroughly mixed. Continue with the rest of the procedure.

1. Rinse and fill 25 mL sample tube (#91980) to 25 mL mark with water to be tested.

NOTE: For results in grains per gallon (gpg), fill to 14.6 mL mark.

2. Add 2 drops R-06380 Phenolphthalein Indicator. Swirl to mix. If sample is colorless, proceed to Step 3. If sample is pink (Fig. 1), add R-06860 Sulfuric Acid N dropwise, swirling after each drop, until color changes from pink to colorless.
3. Add 5 drops R-0630 Chromate Indicator. Swirl to mix. Sample will turn yellow (Fig. 2).
4. Add R-0807 Silver Nitrate Reagent dropwise, swirling and counting after each drop, until color changes from yellow to a milky salmon (brick red) (Fig. 3).

NOTE: A white precipitate will form as R-0807 Silver Nitrate Reagent is added to the sample. Do not add enough R-0807 Silver Nitrate Reagent to give a brown color. First change from yellow to a milky salmon (brick red) is the endpoint.

5. Multiply drops of R-0807 Silver Nitrate Reagent by 2. Record as parts per million (ppm) chloride (Cl^-).

NOTE: For 14.6 mL sample, multiply drops of R-0807 Silver Nitrate Reagent by 0.2. Record as grains per gallon (gpg) chloride (Cl^-).



Fig. 1



Fig. 2



Fig. 3



31 Loveton Circle, Sparks, MD 21152 USA
800-TEST KIT (837-8548) • 410-472-4340

102317