

# COLOR COMPARISON TEST

## pH (3.0-11.0)

Instr. #5067pH

### COMPONENTS:

1 x 5067pH	Instruction
1 x 5425	Color Card Comparator/Instruction
1 x 9017	Test Cell, Calibrated (5 mL) w/ cap, plastic
1 x R-1003U-C-DB	Long Range Indicator, 2 oz, DB

TO ORDER REPLACEMENT PARTS AND REAGENTS CALL TOLL-FREE  
800-TEST KIT (800-837-8548).

### PROCEDURE:

CAREFULLY READ AND FOLLOW PRECAUTIONS ON REAGENT LABELS.  
KEEP REAGENTS AWAY FROM CHILDREN.

NOTE: When dispensing reagents from dropper bottles, **always** hold bottle in a vertical position.

### pH Test

1. Rinse and fill 5 mL test cell (#9017) to 5 mL mark with water to be tested.
2. Add 5 drops R-1003U Long Range Indicator. Cap and mix.
3. Place test cell on white area of Color Card Comparator (#5425). Match color in test cell with a color standard. Record as pH units.



31 Loveton Circle, Sparks, MD 21152 USA  
800-TEST KIT (837-8548) • 410-472-4340