Instr. #5073

DROP TEST FREE CHLORINE (1–10 ppm) & pH (7.0–8.0)

COMPONENTS:

1 x 5073 Instruction 1 x 9056 Comparato

Comparator, Chlorine (hi range), DPD/pH, 7.0-8.0

1 x R-0001-C DPD Reagent #1, 2 oz, Dropper Bottle 1 x R-0002-C DPD Reagent #2, 2 oz, Dropper Bottle

1 x R-0004-C pH Indicator Solution, Phenol Red, 2 oz, Dropper Bottle

TO ORDER REPLACEMENT PARTS AND REAGENTS CALL TOLL-FREE 800-TEST KIT (800-837-8548).

PROCEDURE:

CAREFULLY READ AND FOLLOW PRECAUTIONS ON REAGENT LABELS.
KEEP REAGENTS AWAY FROM CHILDREN.

NOTE: When dispensing reagents from dropper bottles, always hold bottle in a vertical position.

Free Chlorine

- 1. Fill small tube to 9 mL mark with sample water.
- 2. Add 5 drops R-0001 and 5 drops R-0002. Cap and invert to mix.
- 3. Match color.* Record as ppm free chlorine (Cl₂).
- *If color is off-scale: Repeat test using 4.5 mL sample diluted to 9 mL mark with tap water. Multiply reading by 2 to obtain approximate sanitizer level.

pН

- 1. Fill large tube to 44 mL mark with sample water.
- 2. Add 5 drops R-0004. Cap and invert to mix.
- 3. Match color. Record as pH units. If color is between two values, pH is average of the two.

