

COLOR COMPARISON TEST TOTAL BROMINE (0.2-3.0 ppm)

Instr. #5109

COMPONENTS:

MIDGET COMPARATORS*:	RANGE:
9035	0.5-3.0 ppm
9041	0.2-3.0 ppm

*Kit may include only one comparator.

REAGENTS:

1 x R-0001-A	DPD Reagent #1, .75 oz, DB
1 x R-0002-A	DPD Reagent #2, .75 oz, DB

APPARATUS:

1 x 3243	Cap, Test Cell (11.5 mL), plastic
1 x 4024	Test Cell, Calibrated (11.5 mL), plastic
1 x 5109	Instruction
1 x 6002	Brush, Test Cell

TO ORDER REPLACEMENT PARTS AND REAGENTS CALL TOLL-FREE
800-TEST KIT (800-837-8548).

PROCEDURE:

CAREFULLY READ AND FOLLOW PRECAUTIONS ON REAGENT LABELS.
KEEP REAGENTS AWAY FROM CHILDREN.

NOTE: When dispensing reagents from dropper bottles, **always** hold bottle in a vertical position.

Total Bromine Test

1. Rinse and fill 11.5 mL test cell (#4024) to 11.5 mL mark with water to be tested.
2. Add 5 drops R-0001 DPD Reagent #1 and 5 drops R-0002 DPD Reagent #2. Cap and mix.
3. Wipe dry and place in comparator WITH FROSTED SIDE FACING OPERATOR.
5. Match color in test cell with a color standard. Record as parts per million (ppm) total bromine (Br_2).



31 Loveton Circle, Sparks, MD 21152 USA
800-TEST KIT (837-8548) • 410-472-4340