

COLOR COMPARISON TEST

IRON (0-2.0 ppm)

COMPONENTS:

2 x 4028	Pipet, Calibrated (0.5 mL) w/ cap, plastic
1 x 5112	Instruction
1 x 6002	Brush, Test Cell
1 x 9051	Midget Comparator, Iron, Tripyridyl-s-triazine, 0-2.0 ppm
1 x R-0851-A	Iron Reagent #1*, .75 oz
2 x R-0852-A	Iron Reagent #2, .75 oz

**TO ORDER REPLACEMENT PARTS AND REAGENTS CALL TOLL-FREE
800-TEST KIT (800-837-8548).**

PROCEDURE:

**CAREFULLY READ AND FOLLOW PRECAUTIONS ON REAGENT LABELS.
KEEP REAGENTS AWAY FROM CHILDREN.**

Iron Test

1. Rinse and fill 11.5 mL test cell (#4024) to 11.5 mL mark with water to be tested.
2. Using a 0.5 mL pipet (#4028), add 0.5 mL R-0851 Iron Reagent #1. Cap and mix.
WAIT 2 MINUTES.
3. Using a separate 0.5 mL pipet, add 1.0 mL (2 x 0.5 mL) R-0852 Iron Reagent #2.
Cap and mix.
4. Wipe dry and place in comparator WITH FROSTED SIDE FACING OPERATOR.
5. Match color in test cell with a color standard. Record as parts per million (ppm) iron (Fe).

*WARNING: Iron Reagent #1 (R-0851) contains 5-10% hydrochloric acid, a corrosive acid.



31 Loveton Circle, Sparks, MD 21152 USA
800-TEST KIT (837-8548) • 410-472-4340