

COLOR COMPARISON TEST pH (6.8-8.2)

COMPONENTS:

1 x 3243	Cap, Test Cell (11.5 mL), plastic
1 x 4024	Test Cell, Calibrated (11.5 mL), plastic
1 x 4030	Pipet, Calibrated (0.5 & 1.0 mL) w/ cap, plastic
1 x 5199	Instruction
1 x 6002	Brush, Test Cell
1 x 9053	Midget Comparator, pH, Phenol Red, 6.8-8.2
1 x R-1003J-C	pH Indicator Solution, Phenol Red, 2 oz

TO ORDER REPLACEMENT PARTS AND REAGENTS CALL TOLL-FREE
800-TEST KIT (800-837-8548).

PROCEDURE:

CAREFULLY READ AND FOLLOW PRECAUTIONS ON REAGENT LABELS.
KEEP REAGENTS AWAY FROM CHILDREN.

pH Test

1. Rinse and fill 11.5 mL test cell (#4024) to 11.5 mL mark with water to be tested.
2. Using 1.0 mL pipet (#4030), add 0.5 mL R-1003J pH Indicator Solution. Cap and mix.
3. Wipe dry and place in comparator WITH FROSTED SIDE FACING OPERATOR.
4. Match color in test cell with a color standard. Record as pH units.

Instr. #5199

NOTE: If sample color is between two values, pH is average of the two.

NOTE: If determining Acid or Base Demand, save sample and proceed as directed on separate instruction.



31 Loveton Circle, Sparks, MD 21152 USA
800-TEST KIT (837-8548) • 410-472-4340

092717