

# Manganese 0.8

Range(s): 0-0.80 ppm Mn



## Procedure

Note: Glassware that has not been properly cleaned may contaminate the sample and affect test results. If metal contamination is suspected, clean glassware thoroughly before use with Nitric Acid 1N (R-0801); then rinse thoroughly with DI Water (R-0833) or sample water.

1. Turn on the Colorimeter.
2. Select a test menu (ALL TESTS, RECENT TESTS, or FAVORITES) containing Manganese 0.8 using ◀▶.
3. Select Manganese 0.8 using ▲▼; then press ENTER Ⓢ.
4. Rinse and fill 25 mm sample cell to 10 mL mark with DI Water (R-0833), or manganese-free water; then cap and set aside. (This will be the blank sample cell.)

5. Rinse and fill a second sample cell to 10 mL mark with sample water. (This will be the sample.)
6. Using the 0.05 g dipper spoon, add 1 level dipper Manganese 0.8 - Reagent A to each cell; then cap and swirl to dissolve powder.
7. Add 0.5 mL Manganese 0.8 - Reagent B to each cell; then swirl to mix.
8. Add 0.5 mL Manganese 0.8 - Reagent C to each cell; then swirl to mix.
9. Add 0.5 mL Manganese 0.8 - Reagent D to each cell; then cap and swirl to mix thoroughly.
10. Insert blank sample cell into sample cell compartment. Align marks per User's Manual.

11. Select ZERO using ◀▶; then press ENTER Ⓢ. Zero will be displayed.
12. Remove blank sample cell from sample cell compartment.
13. Insert sample cell into sample cell compartment. Align marks.
14. Select READ using ◀▶; then press ENTER Ⓢ. The instrument will read the sample and the result will be displayed.

## Interferences

Alkalinity, Total ( $\text{CaCO}_3$ )  $\geq 300$  ppm – positive interference  
 To remove interference: Fill dilution vial to 50 mL mark and adjust pH to 7-7.5 with Sulfuric Acid N (R-0686).  
 Take a 10 mL portion and follow test procedure above.

Biguanide (as product)  $\geq 250$  ppm – positive interference  
 Bromine  $\geq 20$  ppm – negative interference

Chloride  $\geq 1000$  ppm – positive interference  
 Chlorine  $\geq 10$  ppm – negative interference  
 Copper  $\geq 5$  ppm – positive interference  
 Cyanuric Acid  $\geq 200$  ppm – positive interference  
 Hardness, Calcium ( $\text{CaCO}_3$ )  $\geq 1000$  ppm – positive interference

Hardness, Magnesium ( $\text{CaCO}_3$ )  $\geq 300$  ppm – positive interference  
 Iron, Ferrous  $\geq 2$  ppm – positive interference  
 Permanganate  $\geq 0.6$  ppm – negative interference

## Test Method

PAN (1-(2-Pyridylazo)-2-naphthol)

Under basic conditions PAN chelates with manganese to produce a rust-colored complex that is proportional to the concentration of manganese in a sample.

**Estimated  
Detection Limit**

0.02 ppm Mn

**Precision**

Using two lots of reagent and a standard solution of 0.400 ppm Mn, an individual analyst obtained a standard deviation with the instrument of  $\pm 0.02$  ppm Mn.

**Application**

Recreational Water

**Ordering Info****Reagent Pack**

K-8034 Manganese 0.8

Formulated for exclusive use with Taylor's TTi® Colorimeter.

**Reagent Pack Components**

R-8034A Manganese 0.8 - Reagent A

R-8034B Manganese 0.8 - Reagent B

R-8034C Manganese 0.8 - Reagent C

R-8034D Manganese 0.8 - Reagent D

**Required Reagents & Accessories**

R-0833 DI Water

**Optional Reagents & Accessories**

R-0686 Sulfuric Acid N

R-0801 Nitric Acid 1N



31 Loveton Circle, Sparks, MD 21152 U.S.A.  
800-TEST KIT (837-8548) • 410-472-4340  
customerservice@taylortechnologies.com